

# *In Vitro* Dissolution Testing of Nicotine Release from Smokeless Tobacco Products

*John H. Miller, Fadi Aldeek, Tim Danielson, Yezdi  
B. Pithawalla, Anthony A. Brown and Celeste  
Wilkinson*

*Tobacco Science Research Conference  
Sept. 15-18th 2019*



**Altria**

Altria Client Services

# Substantial Equivalence (SE) Pathway “905”

## Compare Predicate vs. New Product

### Demonstrates if the new product has:

- Same characteristics as the predicate
- Product does not raise different questions of public health

US FDA: Abbreviated HPHC List\*

Cigarette Smoke	Cigarette Filler
NNK	NNK
NNN	NNN
Nicotine	Nicotine
Acetaldehyde	Arsenic
Crotonaldehyde	Cadmium
Formaldehyde	Ammonia
Acrolein	
Acrylonitrile	
Benzene	
1,3-Butadiene	
Isoprene	
Toluene	
4-Aminobiphenyl	
1-Aminonaphthalene	
2-Aminonaphthalene	
Ammonia	
Benzo[a]pyrene	
Carbon monoxide	
	Smokeless Tobacco
	NNK
	NNN
	Nicotine (total and free)
	Acetaldehyde
	Crotonaldehyde
	Formaldehyde
	Arsenic
	Cadmium
	Benzo[a]pyrene

\*Reporting Harmful and Potentially Harmful Constituents in Tobacco Products and Tobacco Smoke Under Section 904(a)(3) of the Federal Food, Drug, and Cosmetic Act” (Guidance for the Industry, March 2012).

# Substantial Equivalence (SE) for Smokable Tobacco Products

## Predicate vs. New Product

- Comparison of HPHCs in cigarette filler and in smoke (route of exposure)
- Smoking machines allow for a comparison of multiple products
- Standardized smoking protocols: ISO, CORESTA & HC
- Not meant to replicate human exposure



**Cigarette**



**Smoking Machine**



**Altria**  
Altria Client Services

ISO: International Organization for Standardization **CORESTA**: Cooperation Center for Scientific Research Relative to Tobacco **HC**: Health Canada.

# Substantial Equivalence (SE) for Smokeless Tobacco Products



**Cigarette**



**Smoking Machine**



**Loose**



**Snus**



**Pouch**



**Altria**  
Altria Client Services

# Dissolution Testing

*In vitro* laboratory test method designed to demonstrate how efficiently an active ingredient is extracted out of a solid oral dosage into solution

## Applications in Pharmaceutical Industry

- Guide product design
- Quality control testing
- Product to product performance comparison
- Develop *in-vivo/in-vitro* correlation (IVIVC)



# Dissolution Methodology Considerations

- Apparatus
- Dissolution Media
- Analytical Method Development & Validation

# Dissolution Apparatus


Multiple commercial & non-commercial options available for measuring release of constituents from oral dosage forms

## Example: U.S. Pharmacopeia Apparatus

Name	Apparatus Type
USP-1	Basket
USP-2	Paddle
USP-3	Reciprocating cylinder
USP-4	Flow-through cell
USP-5	Paddle over disk
USP-6	Cylinder
USP-7	Reciprocating holder

# ALCS Dissolution Method

## SOTAX CE7 Smart Flow-Through Cell Apparatus (USP-4)



- Flow rate 4.0 mL/min

- Temperature 37°C



**Altria**  
Altria Client Services




# ALCS Dissolution Method

## Flow-Through Cell in USP-4 Apparatus


1 g of Loose Tobacco

Filter Chamber →  
3 mm Beads →  
Loose Tobacco →  
1 mm Beads →  
5 mm Red Ruby Bead →



1 Pouch (Snus)

3 mm Beads →  
1 mm Beads →



**Altria**  
Altria Client Services

# ALCS Dissolution Method

## Artificial Saliva Composition\*

Ingredient	Per 1000 mL
Magnesium Chloride Hexahydrate ( $\text{MgCl} \cdot 6\text{H}_2\text{O}$ )	0.17 g
Potassium Hydrogen Phosphate anhydrous ( $\text{K}_2\text{HPO}_4 \cdot \text{H}_2\text{O}$ )	0.68 g
Sodium Chloride ( $\text{NaCl}$ )	0.33 g
Potassium Chloride ( $\text{KCl}$ )	0.75 g
Calcium chloride dihydrate ( $\text{CaCl} \cdot 2\text{H}_2\text{O}$ )	0.15 g
Potassium Carbonate ( $\text{K}_2\text{CO}_3$ )	0.53 g
Type I Water (De-ionized)	1000 mL
Hydrochloric acid	To pH $6.8 \pm 0.1$

\*German Institute for Standardization (DIN) Recipe is based upon German standard DIN v53160-1, "Determination of the Colour Release of Articles of Daily Use, Part1: Resistance to Artificial Saliva", section 4.2, October 2002.



# ALCS Dissolution Method

## Fraction Collection Protocol

Fraction Number	Fraction Collection Time (min)	Fraction Collection Duration (min)	Volume Collected (mL)
1	4	4	16
2	8	4	16
3	12	4	16
4	16	4	16
5	20	4	16
6	30	10	40
7	40	10	40
8	50	10	40
9	60	10	40

- Flow rate 4.0 mL/min
- Temperature 37°C

# Analytical Method: Determination of Nicotine in Dissolution Fractions

## Sample Preparation in Autosampler Vials

1. 0.1 mL of the dissolution fraction
2. 0.8 mL of artificial saliva
3. 0.1 mL of ISTD (Ethyl Benzoate, 1 mg/mL)
4. Cap, vortex and analyze by UPLC-PDA\*

**ISTD:** Internal Standard

**UPLC-PDA:** Ultra-high Performance Liquid Chromatography Photodiode Array

# Analytical Method: Determination of Nicotine in Dissolution Fractions

**Instrument:** Waters Acquity I-Class UPLC system coupled to Photodiode Array Detector

**Column:** BEH C18, 2.1 x 100 mm, 1.7  $\mu$ m, Waters Corporation

**Guard column:** BEH C18 VanGuard, 2.1 x 5 mm, 1.7  $\mu$ m, Waters Corporation

UPLC Parameter		Setting
Run Time		12 min
Injection Volume		10 $\mu$ L
Autosampler Temperature		5 °C
Column Temperature		45 °C
Mobile Phase A	10 mM Ammonium Acetate pH 10	
Mobile Phase B	Acetonitrile	
Pump Program	Gradient Elution	
Flow Rate	0.5 mL /min	



  

Compound	PDA Setting
Nicotine	260 nm
Ethyl benzoate	260 nm



# Analytical Method: Determination of Nicotine in Dissolution Fractions

## UPLC-PDA Chromatograms of Nicotine and Ethyl Benzoate (ISTD)



**ISTD:** Internal Standard

**UPLC-PDA:** Ultra-high Performance Liquid Chromatography Photodiode Array

# Analytical Method: Determination of Nicotine in Dissolution Fractions

## Method Validation

Parameter	Outcome
Calibration (0.5-100 µg/mL)	<ul style="list-style-type: none"><li>▪ <math>R^2 &gt; 0.998</math> on all days</li><li>▪ %RCR &lt; 10%</li></ul>
Accuracy - 3 fortification levels in triplicate for each product type	96.2% - 102%
Repeatability (Intra-day precision, n = 6)	< 2.0%
Intermediate Precision (Inter-days precision, n = 18)	< 6.0%
Specificity	No interferences observed at the retention time of nicotine or IS
LOQ	0.5 µg/mL
Stability - Dissolution samples and final extracts	Stable for up to 15 days when stored in amber glass vials at 0-4°C (refrigerated)



# Analysis of Products for Regulatory Submissions

## Predicate vs. New Product

12 replicates per product\*

- Requires two runs for each product (7 cell holders per unit)
  1. 6 Reps + 1 IQCS (CRP 2.1, American style loose moist snuff)
  2. 6 Reps + 1 reagent blank
- 9 fractions collected for each sample replicate

*\*Guidance for Industry: Dissolution Testing of Immediate Release Solid Oral Dosage Forms, FDA Center for Drug Evaluation and Research (CDER), 1997.*

- **IQCS:** Internal Quality Control Sample







**Altria**

Altria Client Services



# Calculations for Cumulative Release Profile



Error Bars  $\pm 1$  S.D.

# CORESTA Smokeless Tobacco Reference Products (CRPs)

CRP 1.1 (Swedish style snus pouch)



CRP 2.1 (American style loose moist snuff)



CRP 4 (loose-leaf chewing tobacco)



CRP 4.1 (chopped loose-leaf chewing tobacco)




**Altria**  
Altria Client Services

<https://www.coresta.org/coresta-smokeless-tobacco-reference-products>


# Results

## CORESTA Smokeless Tobacco Reference Products (CRPs)



# Results

## CORESTA Smokeless Tobacco Reference Products (CRPs)



# Dissolution Profiles Comparison

Approach based on FDA guidance from CDER\*

- Determine nicotine release profiles for predicate and new product (12 replicates per product)
- Use mean dissolution values from both curves at each time point to calculate difference factor ( $f_1$ ) and similarity factor ( $f_2$ )

$f_1$  values up to 15 (0-15) and  $f_2$  values of 50 or greater (50-100)  
ensure similarity or equivalence between two products

*\*Guidance for Industry: Dissolution Testing of Immediate Release Solid Oral Dosage Forms, FDA Center for Drug Evaluation and Research (CDER), 1997.*



**Altria**  
Altria Client Services

# Results

## Product to Product Comparison using $f_1$ and $f_2$


Compared Products	$f_1$	$f_2$	Equivalency
CRP 2.1 vs CRP 4.1	11.3	52.7	Yes
CRP 2.1 vs CRP 4	27.5	30.0	No
CRP 2.1 vs CRP 1.1	39.2	21.9	No
CRP 4.1 vs CRP 4	23.1	35.3	No
CRP 4.1 vs CRP 1.1	36.1	24.8	No
CRP 4 vs CRP 1.1	20.2	42.2	No

CRP 1.1 (Swedish style snus pouch)  
CRP 2.1 (American style loose moist snuff)  
CRP 4 (loose-leaf chewing tobacco)  
CRP 4.1 (chopped loose-leaf chewing tobacco).

Equivalent:  
f1 values up to 15 (0-15)  
f2 values of 50 or greater (50-100)

# Results


## Commercial Products with Different Flavors



Error Bars  $\pm 1$  S.D.

# Results

## Loose vs Pouched Commercial Products



Error Bars  $\pm 1$  S.D.



# Summary

- An efficient method for the dissolution testing and analytical quantitation of nicotine was validated for a variety of smokeless tobacco products using a USP-4 flow-through cell apparatus and UPLC-PDA
- Percent of total release profiles, and  $f_1$  and  $f_2$  calculations were used to evaluate the similarity and differences between the smokeless tobacco products
- The nicotine release profiles are dependent on the form and cut of the smokeless tobacco products
- Validated dissolution methodologies can be an important tool for smokeless tobacco product assessments and product-to-product comparisons

This presentation may be accessed @  
[www.altria.com/ALCS-Science](http://www.altria.com/ALCS-Science)



# Results

## $f_1$ and $f_2$ Calculations

$$f_1 = \left\{ \left[ \sum_{t=1}^n |R - T| \right] / \left[ \sum_{t=1}^n R \right] \right\} \times 100$$

$$f_2 = 50 \cdot \log \left[ \frac{100}{\sqrt{1 + \frac{\sum_{t=1}^n [R_t - T_t]^2}{n}}} \right]$$

$R_t$  and  $T_t$  are the cumulative percentage dissolved at each of the selected  $n$  time points of the two products.

$f_1$  values up to 15 (0-15) and  $f_2$  values of 50 or greater (50-100) ensure similarity or equivalence between two products.